

1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE
HAZARDOUS (CLASSIFIED) LOCATION
ELECTRICAL EQUIPMENT

This certificate is issued for the following company:

Raytek Corp.
1201 Shaffer Road
Santa Cruz, CA 95060

For:

MX2-NI, MX4-NI RAYNGER MX™ Series Infrared Noncontact Thermometer
NI/II/2/ABCD/T4 Ta = 50°C; NI/II/2/IIC/T4 Ta = 50°C
Special Condition of Use

1. Must be used only with 1.5Volt alkaline batteries
2. Batteries must be changed only in non-hazardous locations
3. Only Raytek temperature probe part # XXXMXTP or XXXMXTCK2 may be connected

Equipment Ratings: Nonincendive for use in Class I, Division 2, Group A, B, C and D and Class I, Zone 2, Group IIC, T4 Ta = 50°C Hazardous indoor (Classified) Locations.

Manufactured By: Raytek GmbH
Blankenburger Strasse 135
D-13127 Berlin, Germany

This certifies that the equipment described has been found to comply with the following Factory Mutual Research Approval Standards:


Approval Standard Class 3600 - 1998
Approval Standard Class 3611 - 1999
Approval Standard Class 3810 - 1989, Supp. No. 1

Approval Job Identification: 3012283 Issue Date: April 19, 2002


Related Approval Job Identification: None

Subsequent Revision Reports/Date Approval Amended: None

Factory Mutual Research



David Styrcula
Technical Team Manager
Approvals Division


Date